



The Johns Hopkins Hospital  
Department of Pathology  
Point-of-Care Testing Office

**NEAR PATIENT TESTING**  
**QUANTIMETRIX DROPPER PLUS URINE CONTROL LOT AND RANGES**  
(Multistix 10SG (Clinitek), Specific Gravity, Pregnancy (OSOM Card Test, hCG CardinalHealth and Clinitest))

**NOTE: Control lot effective 07-15-2025**

Level 1 Lot: **245771**

Level 2 Lot: **245772**

Expiration: 1-31-2026

**Urine Dipsticks Expected Results\*:**

\*Acceptable results determined by Clinitek instrument. If any results are outside of the expected range, the instrument will flag QC result and require corrective action.

**Pregnancy / Specific Gravity by Refractometer Expected Results:**

	Pregnancy	Specific Gravity
Level 1	Negative	1.018 – 1.024
Level 2	Positive	1.012 – 1.018

**ANY RESULT THAT DOES NOT FALL WITHIN THE EXPECTED RESULTS RANGE MUST BE REPEATED AND CORRECTIVE ACTIONS DOCUMENTED ON THE APPROPRIATE QC LOG SHEET.**

**Technical Update (239-19-44-01) Issued 08/27/2019:**

*Endogenous crystalline sediment may be observed in the dropper tip. This does not affect the performance of the product.*