



## POCT Temperature Log

Year: \_\_\_\_\_

Unit/Room: \_\_\_\_\_

Month: \_\_\_\_\_

DAY	Minimum ≥15°C	Maximum ≤30°C	Current Temp. (15°-30°C)	Initials	Comments/Actions Taken	Unit Review	Lab Review
1							
2							
3							
4							
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
<b>NOTE: If outside of Temp. range, note actions taken in comment section and contact POCT Office. (5-2645)</b>							

Revision: 11/22/05

**Comment/Action Codes:**

1. Ensured Room air intake/supply ducts not blocked
2. Adjusted thermostat
3. Continue to monitor
4. Called Facilities/Maintenance- supplies moved to: \_\_\_\_\_ (room location)