



The Johns Hopkins Hospital
 Department of Pathology
 Point of Care Testing Office

NEAR PATIENT TESTING

QUANTIMETRIX DROPPER PLUS URINE CONTROL LOT AND RANGES

(Multistix 10SG (Manual and Clinitek), Specific Gravity, Pregnancy (OSOM Card Test, and Clinitek))

NOTE: Control lot effective August 27, 2021

Level 1 Lot: **135851**

Level 2 Lot: **135852**

Expiration: **08/31/2022**

Urine Dipsticks Expected Results:

	Glucose	Bilirubin*	Ketones	Sp. Gravity	Blood	pH	Protein	Urobilinogen*	Nitrite	Leukocytes
Level 1	Negative	Negative	Negative	1.010 – ≥ 1.030	Negative	5.0 – 6.5	Negative	Normal* (0.2)	Negative	Negative
Level 2	Trace – 2+	1+ - 3+	Trace - 4+	1.010 – 1.025	1+ - 3+	7.5 - ≥9.0	1+ - 3+	2.0 - ≥8.0 EU/dL	Positive	Trace – 3+

*Colors produced by the *urobilinogen* and/or *bilirubin* reactions on Multistix 10SG dipsticks with the urinalysis dipstick control may not be characteristic of those shown on the manufacturer’s label when reading the dipstick reactions visually. *Urobilinogen* reactions are consistent and intensify with the increase in the urobilinogen concentration, but may not provide an exact color match to those displayed on the label.

Pregnancy / Specific Gravity by Refractometer Expected Results:

	Pregnancy	Specific Gravity
Level 1	Negative	1.020 – 1.026
Level 2	Positive	1.012 – 1.018

ANY RESULT THAT DOES NOT FALL WITHIN THE EXPECTED RESULTS RANGE MUST BE REPEATED AND CORRECTIVE ACTIONS DOCUMENTED ON THE APPROPRIATE QC LOG SHEET.

Technical Update (239-19-44-01) Issued 08/27/2019:

Endogenous crystalline sediment may be observed in the dropper tip. This does not affect the performance of the product.